VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N _____ 130195 _____

Supervisor/Operator Performing the Verification Procedure:

A

Name _____ Robert J. Ely _____ ID# ___ 2681 ___ Date ___ 04-15-2009 ___

Agency _____ Craig Police Department _____ Phone # _____ (907) 826-3330 _____

Instrument Location _____ 506 2nd Street, Craig Alaska 99921 _____

Alco S/N ___ 77286 ___ Target Value ___ .079 ___ High Pressure ___ 350 PSI ___

Alco Test Value Average ___ .083 ___ ___ .083 ___
1st Alco 2nd Alco

Signature _____ *Robert J. Ely* _____

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
(3) I am the Scientific Director of the State Breath Alcohol Program.
(4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
(5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 19th day of May, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)

Carolyn M. Noland
NOTARY PUBLIC
STATE OF ALASKA

VERIFICATION OF CALIBRATION REPORT
of DataMaster cdm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N _____ 130195 _____

(CONTINUED FROM FRONT PAGE)

C   TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

APRIL 15, 2009 ✓

OPERATOR'S NAME:
  ELY, RJ
OPERATOR'S NUMBER: 2681
SUBJECT'S LAST NAME:
  VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
  VER OF CAL
D.L. #: 123456789
DEPT/AGENCY: CR01
CASE/REPORT: 2009-XXXX
TEST TYPE: V
ALCO TARGET VALUE: .079 ✓
ALCO S/N: 77286 ✓

   — BREATH ANALYSIS —

.079 ADJUSTED FOR 30.29 in
ALCO TARGET   .079   11:04
BLANK TEST   .000   11:14
INTERNAL STANDARD   VERIFIED   11:15
ALCO TV 30.29 in   .083 ✓   11:15
BLANK TEST   .000   11:16
SUBJECT SAMPLE   .000   11:16
BLANK TEST   .000 ✓   11:17
ALCO TV 30.29 in   .083 ✓   11:17
BLANK TEST   .000   11:18

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

APRIL 15, 2009 ✓
TIME 11:11

   — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c

BAROMETER: 30.29 in

FLOW DETECTOR: OKAY

PUMP
PUMP SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

     PRINTER TEST
!"#&'()*+,-./0123456789:;<=>?@
ABCDEFGHIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghi
pqrstuvwxyz{|}~